### 4NH3-M Ammonia Electrochemical Sensor

(P/N: SEC-4NH3-M)

### **Technical Specifications**

**MEASUREMENT** 

Operating3-electrodePrincipleelectrochemicalDetection Range0~1000 ppmSensitivity8 ± 4nA /ppmResponse Time≤50s (typical 20S)

(T90)

Repeatability <±2% signal Linearity Linear

Long term

<2% signal/month
output Drift</pre>

**ELECTRICAL** 

Resolution 15 ppm

Recommended

Load

5~100 Ω

Bias Voltage 0 mV

**ENVIRONMENTAL** 

Operating -20°C ~ 50°C

Temp. Range

Operating
Humidity
non-condensing

Range

Operating

Pressure 800 ~ 1200 mbar

Range

**LIFETIME** 

Recommended 0°C to +20°C in sealed

Storage Temp. container

Expected

Operating Life 24 months in air

. .

6 months in original

Storage Life packaging

**AERI** 

# JingZhou Aeritech Co.,Ltd.


Tel +86 18995851100, Fax +86 0716 8499894 www.aerisensor.com, E-mail info@aeritech.cn


**Standard** 12 months from date of

Warranty despatch

## **Product Dimensions**









All dimensions in millimeters (± 0.1mm)

# **Temperature Data**

#### **Baseline drift**



#### **Sensitivity Temperature Dependence**



### **Cross-sensitivity Data**

Gas	Concentration (ppm)	Output signal (ppm NH3 equivalent)
Carbon Monoxide	1000	1150
Hydrogen	1000	155
Ethanol	1000	0
Carbon Dioxide	1000	0
Nitrogen Dioxide	50	-100
Hydrogen Sulfide	10	50
Chlorine	20	-45

Whilst the Gas Sensor are designed to be highly specific to the gas they are intended to measure, they will still respond to some degree to various gases. The table below is not exclusive and other gases not included in the table may still cause a sensor to react. The cross-sensitivity values quoted are based on tests conducted on a small number of sensors. They are intended to indicate sensor response to gases other than the target gas. Sensors may behave differently with changes in ambient conditions and any batch may show significant variation from the values quoted.

# **SAFETY NOTE:**

Connection should be made via a PCB mounting socket. Soldering to pins will void the sensor's warranty.

It is important that exposure to high concentrations of solvent vapours is avoided, both during storage, fitting into instruments, and operation;

If the Gas Sensor is removed from application circuit, a jumper should be added on 'R' and 'S' pin.

As applications of use are outside our control, the information provided is given without legal responsibility. Customers should test under their own conditions, to ensure that the sensors are suitable for their own. The data is given for guidance only. It does not constitute a specification or an offer for sale.

