4NO2-S Nitrogen Dioxide Electrochemical Sensor

(P/N: SEC-4NO2-S)

Technical Specifications

Storage Life

Standard

Warranty

6 months in original

18 months from date of

packaging

despatch

Product Dimensions

MEASUREMENT

2 plastrad

Operating 3-electrode
Principle electrochemical

Detection Range 0~20 ppm

Maximum

Overload 250ppm

Sensitivity $0.6 \pm 0.15 \,\mu\text{A /ppm}$ Response Time $\leq 30 \,\text{s(typical 20s)}$

(T90)

Linearity

Repeatability <±2% signal

MEASUREMENT <2% signal/month

Linear

ELECTRICAL

Resolution 0.1 ppm

Recommende

5~30 Ω

d Load

Bias Voltage 0mV

ENVIRONMENTAL

Operating _ -20°C ~ 50°C

Temp. Range

Operating
15% RH ~ 90% RH non-

condensing

Range Operating

Pressure 800 ~ 1200 mbar

Range

All dimensions in millimeters (± 0.1mm)

LIFETIME

Recommende

0°C to +20°C in sealed

d Storage

container

Expected

Temp.


Operating 24 months in air

Life




Tel +86 18995851100, Fax +86 0716 8499894 www.aerisensor.com, E-mail info@aeritech.cn






Temperature Data

Baseline drift



Sensitivity Temperature Dependence





Cross-sensitivity Data

Gas	Concentration (ppm)	Output signal (ppm NO2 equivalent)
Carbon Monoxide	200	0
Nitric Oxide	30	0
Sulphur Dioxide	5	-5
Hydrogen Sulfide	15	-0.1

Whilst the Gas Sensor are designed to be highly specific to the gas they are intended to measure, they will still respond to some degree to various gases. The table below is not exclusive and other gases not included in the table may still cause a sensor to react. The cross-sensitivity values quoted are based on tests conducted on a small number of sensors. They are intended to indicate sensor response to gases other than the target gas. Sensors may behave differently with changes in ambient conditions and any batch may show significant variation from the values quoted.

SAFETY NOTE:

Connection should be made via a PCB mounting socket. Soldering to pins will void the sensor's warranty.

It is important that exposure to high concentrations of solvent vapours is avoided, both during storage, fitting into instruments, and operation;

If the Gas Sensor is removed from application circuit, a jumper should be added on 'R' and 'S' pin.

As applications of use are outside our control, the information provided is given without legal responsibility. Customers should test under their own conditions, to ensure that the sensors are suitable for their own. The data is given for guidance only. It does not constitute a specification or an offer for sale.

