4PH3-L Phosphine Electrochemical Sensor

(P/N: SEC-4PH3-L)

Technical Specifications

packaging

Product Dimensions

MEASUREMENT

Standard 18 months from date of

Warranty

despatch 3-electrode

Principle electrochemical 0~1000 ppm **Detection Range**

Maximum

 $0.07 \pm 0.0.05 \mu A / ppm$ **Overload**

Sensitivity ≤60s

Response Time

<±2% signal (T90)

Repeatability Linear

Linearity <2% signal/month

ELECTRICAL

d Load

Operating

Resolution 1 ppm

Recommende

5~30 Ω

Bias Voltage 0mV

ENVIRONMENTAL

Operating -20°C ~ 50°C

Temp. Range

Operating 15% RH ~ 90% RH non-


Humidity

condensing Range


Operating

Pressure 800 ~ 1200 mbar

Range



DOES NOT INCLUDE THIKNESS OF LABEL ON LABELLED VARIANTS



All dimensions in millimeters (± 0.1mm)

LIFETIME

Recommende

0°C to +20°C in sealed d Storage

container Temp.

Expected

24 months in air **Operating**

Life

Storage Life 6 months in original




JingZhou Aeritech Co.,Ltd.


Tel +86 18995851100, Fax +86 0716 8499894 www.aerisensor.com, E-mail info@aeritech.cn

Temperature Data

Baseline drift



Sensitivity Temperature Dependence





Cross-sensitivity Data

Gas	Concentration (ppm)	Output signal (ppm PH3 equivalent)
Carbon Monoxide	200	0
Nitric Oxide	30	0.5
Nitrogen Dioxide	5	-1
Hydrogen Sulfide	15	6

Whilst the Gas Sensor are designed to be highly specific to the gas they are intended to measure, they will still respond to some degree to various gases. The table below is not exclusive and other gases not included in the table may still cause a sensor to react. The cross-sensitivity values quoted are based on tests conducted on a small number of sensors. They are intended to indicate sensor response to gases other than the target gas. Sensors may behave differently with changes in ambient conditions and any batch may show significant variation from the values quoted.

SAFETY NOTE:

Connection should be made via a PCB mounting socket. Soldering to pins will void the sensor's warranty.

It is important that exposure to high concentrations of solvent vapours is avoided, both during storage, fitting into instruments, and operation;

If the Gas Sensor is removed from application circuit, a jumper should be added on 'R' and 'S' pin.

As applications of use are outside our control, the information provided is given without legal responsibility. Customers should test under their own conditions, to ensure that the sensors are suitable for their own. The data is given for guidance only. It does not constitute a specification or an offer for sale.

